


SABRE Medical Oxygen Regulators

Sabre Medical Oxygen Regulators must be serviced **every five years**. Details for the next date of service can be found on the waist label. (see the diagram below). Cylinder Test Period and Regulator Service must be kept in line.

Example



**PLEASE CHECK THE SERVICE DUE DATE
ON YOUR REGULATOR**